

Appendix 10. Certified Test Report



CERTIFIED TEST REPORT

CUSTOMER: MCCALLS METER SALES SERVICE
MODEL NO: M0306
METER SERIAL NO: 11-04485

CONFIGURATION

METER INSIDE DIAMETER: 6.065
METER OUTSIDE DIAMETER: 6.625
TEST DATE: 6/16/2011
TEST FACILITY: Volumetric
IDEAL TEST CONSTANT: 6738

CALIBRATION DATA

| | Tested TC | GPM | Accuracy |
|---|-----------|------|----------|
| 1 | 6741 | 1257 | 100.0 |

CERTIFIED BY: Paul Hobbs DATE: 6/20/2011

This calibration was performed on a gravimetric or volumetric test facility, traceable to the National Institute of Standards and Technology, USA. The estimated flow measurement uncertainty of the calibration facilities are:
Gravimetric +/- 0.15% Volumetric +/- 0.5%



McCROMETER

3255 WEST STETSON AVENUE

HEMET, CA 92546 USA

PHONE (951) 652-6811 / FAX (951) 652-3078

WEB SITE: <http://www.mccrometer.com> E-MAIL: info@mccrometer.com



11-04485

6/20/2011 9:06:38 AM
Version 1.2 (4/18/2007)